

# Supplies for Agilent Instruments: Inlet Liners

Splitless Liners for Agilent GCs	Benefits/Uses	ID* / OD & Length (mm)	Similar to Agilent part #	ea.	cat.# 5-pk.	25-pk.
	trace samples <2µL	2.0 ID 6.5 OD x 78.5	5181-8818 (ea.) 5183-4703 (5-pk.) 5183-4704 (25-pk.)	20712	20713	20714
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	210-3003 (ea.) 210-3003-5 (5-pk.)	20772	20773	20774
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	—	20772-214.1	20773-214.5	20774-214.25
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	19251-60540 (ea.) 5183-4691 (5-pk.) 5183-4692 (25-pk.)	22400	22401	22402
	trace samples <2µL	2.0 ID 6.5 OD x 78.5	18740-80220 (ea.) 5183-4707 (5-pk.)	20914	20915	—
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	—	20912	20913	—
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	—	22403	22404	—
	trace samples <2µL	2.0 ID 6.5 OD x 78.5	—	20795	20796	20797
	trace samples <2µL	2.0 ID 6.5 OD x 78.5	—	20795-214.1	20796-214.5	20797-214.25
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	5181-3316 (ea.) 5183-4695 (5-pk.) 5183-4696 (25-pk.)	20798	20799	20800
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	—	20798-214.1	20799-214.5	20800-214.25
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	5062-3587 (ea.) 5183-4693 (5-pk.) 5183-4694 (25-pk.)	22405	22406	22407
	trace samples >2µL	4.0 ID 6.5 OD x 78.5	—	22405-213.1	22406-213.5	22407-213.25
	trace, active samples >2µL	4.0 ID 6.5 OD x 78.5	5181-3315 (ea.) 5183-4705 (5-pk.) 5183-4706 (25-pk.)	20784	20785	20786
	trace, active samples >2µL	4.0 ID 6.5 OD x 78.5	—	20784-214.1	20785-214.5	20786-214.25
	trace, active, dirty samples <2µL	2.0 ID 6.5 OD x 78.5	—	20907	20908	—
	trace, active, dirty samples >2µL	4.0 ID 6.5 OD x 78.5	—	20895	20896	20997
	trace, active, dirty samples >2µL	4.0 ID 6.5 OD x 78.5	—	20895-214.1	20896-214.5	20997-214.25
	base easily packs with wool for dirty samples <2µL	2.0 ID 6.5 OD x 78.5	—	20980	20981	20982
	base easily packs with wool for dirty samples >2µL	4.0 ID 6.5 OD x 78.5	—	20983	20984	20985
	base easily packs with wool for dirty samples >2µL	4.0 ID 6.5 OD x 78.5	—	20983-214.1	20984-214.5	20985-214.25
	base easily packs with wool for dirty samples >2µL	4.0 ID 6.5 OD x 78.5	—	22408	22409	22410
	trace, active samples >2µL	4.0 ID 6.5 OD x 78.5	—	20986	20987	20988

\*Nominal ID at syringe needle expulsion point.

\*\*Use with two-hole ferrule for dual-column analysis.

†Use this liner for increased sensitivity.

 All liners are  
**100%**  
 deactivated

 All liners are shipped intermediate  
 polarity (IP) deactivated unless  
 otherwise requested.